

Contents

Editorial: Going from strength to strength	1
Harmonisation of conductivity tests for Pharmaceutical Waters <i>AC Bevilacqua</i>	3
Pharmaceutical isolator leak testing – proposal for specification of leak test rates which allows a comparison of quoted values <i>SA Burnett</i>	13
Freeze-drying – a review <i>K Murgatroyd</i>	21
Scanning electron and atomic force microscopy for the sizing of bacterial populations <i>LM Madsen</i>	27
SHORT COMMUNICATION The use of modular concepts in the construction of a parenteral pharmaceutical facility <i>RJ Meyer</i>	31
Dates for your diary	35
Book review	37
Index to Volume 5	38

Instructions for authors in this issue

Contents and Abstract

Harmonisation of conductivity tests for Pharmaceutical Waters

Anthony C Bevilacqua, PhD
*Research & Sensor Development Manager, Thornton Inc., Waltham, MA,
USA*

The chemistry tests to verify the quality of bulk Pharmaceutical Waters underwent substantial changes with the implementation of USP 23 in 1996. These qualitative tests were replaced with modern, quantitative, instrument-based tests such as conductivity and total organic carbon (TOC). The acceptance of these changes was not without its problems, but the resulting changes have improved Pharmaceutical Water quality control. Subsequently, the European Pharmacopoeia (EP) implemented its versions of conductivity and TOC tests for Pharmaceutical Waters. The conversion to these instrument-based tests represented a significant step towards unifying the tests for these EP and USP waters. The USP TOC monograph (<643>) is substantially identical to its operation and implementation in the EP (2.2.44). Interestingly, the less complicated and more technologically robust method of conductivity greatly differs between the USP and EP. The reasons for these differences are based on different application strategies. The similarities in these strategies are discussed. However, the practices espoused in the EP standard possess some technical faults or uncertainties which require reconsideration based on the science behind the water chemistry and conductivity.

Part of this paper was presented at The Parenteral Society meeting on September 14, 1999 in London, UK.

Corresponding author: Anthony C. Bevilacqua, PhD, Research & Sensor Development Manager, Thornton Inc., 1432 Main Street, Waltham, MA 02451-1623, USA. Tel: +001 (781) 890 3399 Fax: +001 (781) 890 5506 Email: abevilacqua@thorntoninc.com

Pharmaceutical isolator leak testing – proposal for specification of leak test rates which allows a comparison of quoted values

SA Burnett, BEng (Hons), MSc
Senior Engineer, Gravatorm Engineering Systems Ltd.

At present, most isolator manufacturers quote leak rates in pressure drop/rise per unit time. This does not, however, allow a true comparison, as the tests are frequently performed at different test pressures, i.e. 1000 Pa above or below atmosphere compared with 250 Pa above or below atmosphere. Furthermore, it gives no idea of leak rate volume per unit time. This is important, as it is difficult to picture a Pascal of pressure leaking, whereas an operator can picture a volume (i.e. the number of cc's), allowing him or her to make a judgement on the contamination that could be going in or out of the containment. The proposal in this paper is to provide a

standardised leak rate in terms of an agreed standard pressure (not necessarily the test pressure) and volume leaked in cc per unit of time (i.e. leak rate = kPa cc/sec).

Corresponding author: Simon A. Burnett, Project Manager – OEM, Gravatom Engineering Systems Ltd., Gravatom Business Centre, Claylands Road, Bishops Waltham SO32 1BH, UK. Tel. +44 (0) 1489 896010, Fax. +44 (0) 1489 894382. Website: www.gravatom.com

Freeze-drying – a review

Kevin Murgatroyd

Pharmaceutical Consultant, Bovis Lend Lease Pharmaceuticals

This paper is a general review of the freeze-drying process and the freeze-dryer, primarily in the pharmaceutical arena. After defining freeze-drying, the rationale for the use of this process is discussed. The salient points of the three phases of freeze-drying are identified. The evolution of the freeze-dryer is described and the GMP implications of this evolution are noted and described. Present and future developments, notably loading systems, are discussed and the two goals of freeze-dryer development are identified. However, it is noted that these two goals are unlikely to find favour in the pharmaceutical marketplace.

Corresponding author: Dr Kevin Murgatroyd, Pharmaceutical Consultant, Bovis Lend Lease Pharmaceuticals, Tanshire House, Shackleford Road, Elstead, Surrey GU8 6LB. Tel: +44 1252 703663, fax: +44 1252 703 684. Email: Kevin.Murgatroyd@eu.bovislendlease.com.

Scanning electron and atomic force microscopy for the sizing of bacterial populations

Madsen LM, Steves MF, Howard, Jr. GW

Pall Corporation, Scientific and Laboratory Services Department, Port Washington, New York, USA

Microporous membrane filtration is routinely used in the laboratory and in pharmaceutical manufacturing plants for the sterilisation of aseptic processed drug products and other solutions. Currently, *Brevundimonas diminuta* is used as a standard challenge organism to validate 0.2/0.22- μ m rated "sterilising grade" filters and most commonly applied in "worst case" (parametric) drug product specific challenges to validate filtration sterilisation processes. However, other organisms that may be more penetrating and

presumptively smaller than *B. diminuta* are often identified as bioburden and therefore sizing of bacteria becomes an important criterion for appropriate membrane filter selection. Most commonly, bacterial populations are sized by scanning electron microscopy (SEM). The average size of a cell in a population of *B. diminuta* ranges from 0.32-0.35 by 0.82-0.96 μm when measured by SEM. The potential effect of preparation methods on size measurements is unresolved. We have also been investigating the use of atomic force microscopy (AFM), a technique that does not require extensive sample manipulation and dehydration, for sizing. Using AFM, our initial observations for the average *B. diminuta* cell size is $0.609 \pm 0.084 \mu\text{m}$ by $1.077 \pm 0.185 \mu\text{m}$. These data suggest that dehydration during sample preparation for SEM may lead to smaller apparent size. Previous work has demonstrated that *Ralstonia pickettii*, a common bioburden organism, can penetrate 0.2- and 0.22- μm rated filters if exposed to certain pharmaceutical process conditions². SEM and AFM sizing data for populations of *R. pickettii* are presented and the relative merits of both techniques are discussed.

Corresponding author: Lisa M. Madsen, Ph.D., Staff Scientist, Pall Corporation, Scientific and Laboratory Services Department, Port Washington, New York, USA Tel: (516) 484-3600 x6332 Fax: (516) 484-3628 email: Lisa_Madsen@pall.com

Short communication

The use of modular concepts in the construction of a parenteral pharmaceutical facility

Ronald J Meyer
Dendron Associates, Inc. (Formerly, Strategic Project Planner, Eli Lilly & Company, retired)

The recent development of a modular approach for delivering a complete pharmaceutical manufacturing facility has resulted in technical and business advantages over conventional construction. A case study of a liquid parenteral formulation and fill facility for Eli Lilly, Egypt is presented here.

Corresponding author: Ronald J Meyer, Denron Associates, Inc. 11300 W. Maple Drive, Columbus, Indiana 47201-4548, USA. Tel: +1 812 342 1484/3444 Fax: +1 812 342 1484. Email: ronmeyer@iquest.net

VOLUME 6 NUMBER 2 2001

Contents

Editorial: Never a dull moment	41
Harmonising the training of Qualified Persons in Europe <i>P Davies</i>	43
Pressure differential testing of sealed vials containing freeze-dried product <i>JW Strike</i>	47
Risk assessment in the manufacture of medicinal products <i>S Lovtrup</i>	53
OPINION PAPER The future of aseptic processing of healthcare products <i>JE Akers</i>	59
SHORT COMMUNICATION Towards a ‘human pyrogen test’ <i>S Poole and RE Gaines Das</i>	63
SHORT COMMUNICATION Integral calculus to evaluate decontamination times in isolators <i>G Mercey</i>	65
Dates for your diary	67
Book review	69

Instructions for authors in Vol 6 No 1

Contents and Abstract

Harmonising the training of Qualified Persons in Europe

A review of the role and status of the Qualified Person (QP) throughout Europe

Dr Peter Davies

Pharmaceutical Training Systems Ltd., East Preston, West Sussex, UK

In this paper, the author considers the different ways in which Qualified Persons (QPs) are trained and appointed in various Member States of the EU. He examines some of the basic differences between The Pharmacien Responsable in France as compared to the QP in the UK. The outline syllabi of pharmacy degrees are discussed in the light of EC Directive educational intentions. The author presents a short review of how a number of Member States train and appoint their QPs. Finally, he suggests possible routes for overcoming the procedural differences between Member States in order to rationalise training and appointment of QPs throughout Europe.

Corresponding author: Dr Peter Davies, Pharmaceutical Training Systems Ltd. 1 The Roystons, East Preston, West Sussex BN16 2TR, GB Tel/fax: +44 (0)1903 78 45 81, e-mail: pharmatraining@roystons.fsnet.co.uk

Pressure differential testing of sealed vials containing freeze-dried product

JW Strike

GlaxoSmithKline Operations, Dartford, Kent, UK

An in-house requirement to carry out 100% inspection of vials of freeze-dried product for leaks resulted in the search for a suitable method that could be validated and that was capable of a high throughput rate. This paper presents findings from the method selection process, and a description of the methods and results of a successful validation exercise on the selected automated Wilcomat R18F Pressure Differential Tester, using a proceduralised dye bath test method as a reference point. A brief explanation of the machine operating principles and summary of operating experience are also presented.

Corresponding author: Mr Jeremy Strike, GlaxoSmithKline Operations, Temple Hill, Dartford, Kent, DA1 5AH, UK. Tel: +44 (0)1322 392 849, Fax: +44 (0)1322 391061, email: jbs63857@glasosmithkline.co.uk

Risk assessment in the manufacture of medicinal products based on Design and Barrier Assessment (DaBA)

Steen Løvtrup

Principal Scientist, Validation, Novo Nordisk, Copenhagen, Denmark

Risk assessment in context of GMP and validation has been discussed for a number of years but not applied to the manufacture of medicinal products in reality. The reason for this may be that established methods are general, rather than being configured to focus on aspects that may adversely affect product quality. In spite of this, there is no doubt that risk assessment will be an important decision tool with regard to criticality; one suitable method could be Design and Barrier Assessment, based on the principles of HACCP and configured to focus on product safety and efficacy.

Corresponding author: Steen Løvtrup, B.Sc. Principal Scientist, Validation
Novo Nordisk, Niels Steensensvej 1, DK 2820 Gentofte, Denmark E-
mail: slq@novonordisk.com Tel: +(45)44439756 Fax: +(45)44438383

Opinion Paper

The future of aseptic processing of healthcare products

James E Akers

President, Akers Kennedy and Associates, Kansas City, MO, USA

There are two processes by which pharmaceutical products labelled sterile are manufactured, aseptic processing and terminal sterilisation. In current regulatory guidance and industry standards, the terms "aseptic" is often used interchangeably with the term "sterile", although scientifically, the words have very different meanings, aseptic being free from pathogens, rather than sterile. Aseptically manufactured products require the exclusion of micro-organisms from the product stream and from the environment in which the product is filled or assembled.

Given the necessity of human operator involvement in aseptic manufacturing in conventional cleanrooms, aseptically produced products have been considered less safe than terminally sterilised products. However, the emergence of new advanced manufacturing technologies that effectively eliminate the likelihood of human-borne contamination in product filling or assembly may usher in an era in which the safety of products manufactured using exclusion technologies is so close to that of sterilised products, that the difference can be considered negligible from a consumer safety perspective. The continued evolution of clean manufacturing environments will have profound effects on product manufacturing regulation and testing requirements within the next decade. This article examines the potential of the new technology and the changes its implementation promises to bring to sterile product manufacturing.

Corresponding author: Dr James E Akers, President, Akers, Kennedy and Associates, 125 West 60th Terrace, PO Box 22562, Kansas City, MO, USA.
Tel: +1 816 822 7444, Fax: +1 816 822 7555, email: AKAINCK@aol.com

Short Communication

Towards a 'human pyrogen test'

Stephen Poole¹ and Rose E Gaines Das²

Divisions of Endocrinology¹ and Informatics², National Institute for Biological Standards and Control, Blanche lane, South Mimms, Potters Bar, Herts, EN6 3QG.

In the twenty years or so since an in vitro pyrogen test based upon the human fever reaction, a 'human (in vitro) pyrogen test', was first considered, a number of variants of this test have been developed and are at various stages of development. These are reviewed below.

Corresponding author: Stephen Poole, Division of Endocrinology, NIBSC, Blanche Lane, South Mimms, Potters Bar, Herts, EN6 3QG. Tel: +44 (0)1707 654753, Fax: +44 (0)1707 646730, e-mail: spoole@nibsc.ac.uk.

Short Communication

Integral calculus to evaluate decontamination times in isolators

Gilles Mercey, GMI,
Vendôme, France

The first rules for calculating decontamination of isolators based on research conducted over a number of years are described. The work sheds some light on the results obtained using LOGI-STER-2SR for decontamination of isolators by means of chemical vapour. Some data are omitted for commercial confidentiality reasons.

Les travaux entrepris par GMI depuis plusieurs années permettent aujourd'hui d'énoncer les premières lois de calcul concernant la décontamination des isolateurs. Certes, pour des raisons de concurrence, tout ne sera pas dit.

G.M.I. donne ici quelques éléments sur ses résultats obtenus avec LOGI-STE-2SR pour la décontamination des isolateurs par vapeur chimique.

Corresponding author: Mr Gilles Mercey, GMI, 7 Tertre de la Motte, 41100
Vendôme, France. Tel/Fax: +33 2 54 77 04 01 and +33 6 85 55 50 36.
Website: www.webtours.fr/gmi

Contents

Editorial: International outreach	71
High hydrostatic pressure (HHP): An alternative method of sterilisation and decontamination of fragile drugs? <i>Y Rigaldie, G Lemagnen, A Largeteau, D Larrouture, M Abba, C Durandeu B Vallayer, L Grislain and G Demazeau</i>	73
Air treatment systems in the pharmaceutical industry Part I: General aspects of design, filtering and air conditioning Part II: Air conditioning in pharmaceutical plants manufacturing non-sterile dosage forms: practical examples <i>RS Macian and IL Riero</i>	79
The bacterial endotoxins test: past, present and future <i>JF Cooper</i>	89
Annex 17 – parametric release: A review from the perspective of a Qualified Person <i>S Green</i>	95
Modulated temperature differential scanning calorimetry and its application to freeze-drying <i>V Kett</i>	99
Dates for your diary	105
Book review	106

Contents and Abstract

High hydrostatic pressure (HHP): An alternative method of sterilisation and decontamination of fragile drugs?

Y. Rigaldie^{1,2,3}, G. Lemagnen¹, A. Largeteau², D. Larrouture¹, M. Abba^{1,3}, C. Durandeu¹, B. Vallayer³, L. Grislain¹, G. Demazeau²

1. *Laboratoire de Technologie Pharmaceutique Industrielle de Bordeaux, Faculté de Pharmacie, Université Victor Segalen Bordeaux 2, Bordeaux, France*
2. *Interface Hautes Pressions, ENSCPB - ICMCB, Talence, France*
3. *Ellipse Pharmaceuticals, Pessac, France*

The aim of this study was to investigate the use of HHP technology as a possible alternative method for the decontamination and sterilisation of fragile drugs. We demonstrated the safety of HHP treatment on several fragile biomolecules (molecular weight from 1200 to 150 000 g/mol: two peptides, insulin and antibodies). A HHP treatment (10 min at 400 MPa) inactivates pure suspensions of *Pseudomonas aeruginosa* and *Candida albicans*. HHP at low temperature induces a stronger inactivation than at room temperature. *Staphylococcus aureus* needs a longer duration of HHP treatment to be totally inactivated. Since the vegetative cells of endospores appear to be barosensitive, further research is being carried out on inducing the germination of *Bacillus subtilis* spores under HHP.

L'objet de ces travaux est d'étudier l'utilisation de la technologie des Hautes Pressions comme une méthode alternative potentielle de décontamination et de stérilisation des molécules fragiles. Nous montrons l'innocuité du traitement HP sur différentes biomolécules fragiles (poids moléculaire croissant de 1200 à 150 000 g/mol : deux peptides, de l'insuline et des anticorps). Un traitement sous hautes pressions (10 min à 400 MPa) permet d'inactiver totalement des suspensions pures de *Pseudomonas aeruginosa* et de *Candida albicans*. Le traitement HP appliqué à basses températures induit une plus forte inactivation qu'à température ambiante. *Staphylococcus aureus* nécessite un temps de traitement plus long pour être inactivé. Comme les formes végétatives des endospores semblent être barosensibles, plusieurs protocoles de germination par HP des spores de *Bacillus subtilis* sont actuellement à l'étude.

Corresponding author: Mr Yohan Rigaldie, Laboratoire de Technologie Pharmaceutique Industrielle de Bordeaux, Faculté e Pharmacie, Université Victor Segalen Bordeaux 2, Bordeaux, France Tel: +33 (0) 5 57 57 12 40 Fax: +33 (0) 5 56 96 05 67, email: yohan.rigaldie@ltp.u-bordeaux2.fr

Air treatment systems in the pharmaceutical industry

Part I: General aspects of design, filtering and air conditioning

Ramón Salazar Macian* and Ignacio Lerín Riera**

** Professor of Pharmaceutical Technology, School of Pharmacy, University of Barcelona, Barcelona, Spain*

*** Dr Ignacio Lerín Riera, is attached to the Pharmacy and Pharmaceutical Technology Department of the University of Barcelona's Pharmacy School, and is the Quality Assurance Director at Glatt Labortechnik, S.A.*

Nowadays, air and water are deemed to be fundamental to the quality of manufactured pharmaceutical products. Thus, air and water treatment systems must be integrated into the relevant industrial manufacturing process and fully validated. These papers (Parts I and II) attempt to bring up to date the air treatment process in a pharmaceutical industrial plant manufacturing non-sterile medicinal products. Part I considers general aspects of design, filtering and air conditioning. In Part II, various examples of clean-room air conditioning are presented.

Hoy día el aire y el agua en la Industria Farmacéutica, se consideran básicos para la calidad de los productos que se fabrican. Por ello los sistemas de tratamiento de aire y agua han de estar integrados en el proceso de fabricación industrial de medicamentos y han de estar validados. El presente trabajo, pretende ser una puesta al día, del tratamiento de aire en una Planta Industrial Farmacéutica de fabricación de medicamentos no estériles. En una primera parte (parte I), se trata de consideraciones generales, de diseño, filtración y climatización. En una segunda parte (parte II) se muestran ejemplos de climatización de salas limpias.

Contact address: Profesor Dr. Ramón Salazar Macian. Unidad de Tecnología Farmacéutica, Departamento de Farmacia y Tecnología Farmacéutica, Universidad de Barcelona, Avenida Joan XXIII, s/n, 0828 Barcelona, España Tel: (+34) 93 403 5861, Fax: (+34) 93 403 5936, email: salazar@farmacia.far.ub.es

The bacterial endotoxins test: past, present and future

Dr James F Cooper

Founder and Scientific Director, Charles River Endosafe, Clinical Professor of Pharmaceutical Sciences, Medical University of South Carolina, Charleston, USA

The bacterial endotoxins test (BET) became the test of choice for detecting pyrogen in parenterals because of its simplicity, sensitivity and specificity. A harmonised BET that simplified and included all LAL methods became effective in 2001. Future revision may include certification of endotoxin standards, tighter acceptance criteria for kinetic LAL methods and sample preparation by concentration. Endotoxin alert limits are proposed for starting

materials to reduce the risk of pyrogenic products. Broad use of the BET in process control has enhanced the microbiological purity of finished products.

Corresponding author: Dr James F Cooper, 1056 Fort Sumter Drive, Charleston, SC 29412, USA. Tel: (+1) 843-795-7316; fax: (+1) 843-795-7221 Email: jimandfran@worldnet.att.net

Modulated temperature differential scanning calorimetry and its application to freeze-drying

Vicky Kett

The School of Pharmacy, The Queen's University of Belfast, Belfast, UK

MTDSC is a software modification of the traditional DSC thermal analysis technique that allows more accurate determination of the glass transition as well as measurement of the endothermic relaxation that often accompanies the transition. The glass transition is an essential parameter both of the original frozen solution and of the end product. Measurement of endothermic relaxation allows the determination of molecular relaxation times in the freeze-dried product that may be useful in predicting the effect of formulation variables and storage conditions on physical stability.

Corresponding author: Vicky Kett, Lecturer in Pharmaceutics, The School of Pharmacy, The Queen's University of Belfast, 97 Lisburn Road, Belfast BT9 7BL. Tel: +44 (0)2890 272008 Fax: +44 (0)2890 272028 email: v.kett@qub.ac.uk

Annex 17 – Parametric release:

A review from the perspective of a

Qualified Person

Stewart I Green

Director of Quality, Wyeth Manufacturing (UK), Havant, Hampshire, UK

This paper evaluates Annex 17 'Parametric Release', published in the EU, which provides the first guidance within Europe as to how manufacturers can reduce finished product testing. It critically examines the removal of sterility testing, and the removal/reduction of finished product testing and raw materials testing. Although such removals or reductions are of great interest to many manufacturers, the steps required to effect such a process and to

achieve regulatory approval are considerable. The paper looks at each situation and examines the requirements and, in particular, their impact on the Qualified Person, who, at the end of the day, is responsible for releasing product produced under such a regime.

Corresponding author: Stewart I Green, Wyeth Manufacturing (UK), New Lane, Havant, Hants PO9 2NG Tel: 023 9268 2258, Fax: 023 9249 8017, email: greens5@labs.wyeth.com

Contents

INTRODUCTION TO THIS SPECIAL ISSUE

Microbiological aspects of parenteral production: recent developments

N Hodges and G Hanlon 111

MICROBIOLOGY PAPERS

Microbiological environmental monitoring in today's industry

SM Johnson 115

Product Testing

R Baird 121

Viabile but non-culturable microorganisms

P Newby 125

Biological indicators – friend or foe?

D Allison, P Gilbert and N Halls 131

Challenges posed by biotechnology

B Pietrowski 135

OTHER PAPER

Leaking aspects of isolators

B Midcalf 139

Letter to the editor

145

Book review

147

Dates for your diary

149

Contents and Abstract

Introduction to this special issue

Microbiological aspects of parenteral production: recent developments

Norman Hodges and Geoff Hanlon, guest editors
School of Pharmacy and Biomolecular Sciences, University of Brighton, UK

Microbiological environmental monitoring in today's industry

Sharon M Johnson
European Microbiology Manager, Baxter Healthcare Ltd., Newbury, Berkshire, UK.

Increasing emphasis has been placed on environmental monitoring over recent years when, at the same time, there have been improvements in general facility design, use of isolator technology and greater focus on operator training. In recognising the limitations of the monitoring techniques available to the microbiologist, which have not changed over time, it is important that the monitoring programme is appropriately targeted and the use of HACCP, as a process management tool, is increasingly topical.

Corresponding author: Sharon M Johnson, European Microbiology Manager, Baxter Healthcare Ltd. Tel: +44 (0)1635 206000; fax: +44 (0)1635 206115; email: sharon_johnson@baxter.com

Product testing

Rosamund Baird
Recent Daphne Jackson Fellow, School of Pharmacy and Pharmacology, University of Bath, UK

The purpose, interdependence and limitations of bioburden testing, sterility testing and parametric release in the manufacture of sterile pharmaceutical products are considered in this article. Test methods are also discussed.

Corresponding author: Rosamund Baird, Lyes House, Hummer, Trent, Sherborne, Dorset DT9 4SH, Tel: 01935 850926, Fax: 01935 851097.

Viable but non culturable (VBNC) microorganisms in the pharmaceutical industry – their significance and detection

Paul Newby
GlaxoSmithKline Global Manufacturing and Supply

The purpose of this paper is to assess the potential significance of Viable but Non Culturable (VBNC) micro-organisms in the pharmaceutical industry. Consideration is given to the nature of the VBNC state. Current methods for the detection of such organisms are outlined and potentially significant new methods, which may impact the industry, are considered. The significance of VBNC organisms on the future direction of pharmaceutical microbiology is discussed.

Corresponding author: Paul Newby, GlaxoSmithKline Global Manufacturing and Supply, Harmire Road, Barnard Castle, Co. Durham DL12 3DT.

Biological Indicators: friends or foes?

1DG Allison, 1P Gilbert and 2NA Halls

1School of Pharmacy and Pharmaceutical Sciences, University of Manchester, Manchester, UK and 2GlaxoSmithKline, Global Manufacturing and Supply, Stockley Park West, Uxbridge, UK.

There is a growing concern in the pharmaceutical and healthcare industries over the selection, use and properties of Biological Indicators (BIs) for the validation and monitoring of sterilisation processes. BIs possess a known sterilisation resistance and are used to assess the microbial lethality of a given sterilisation process. Collectively, they are well-characterised populations of bacterial endospores, their selection being governed principally by their high resistance to specific sterilisation processes in relation to the indigenous microbial bioburden. Specific strains are used for the validation of steam, dry heat, radiation and ethylene oxide and monitoring of ethylene oxide sterilisation processes. In theory, therefore, BIs provide a well-defined, reliable but crude method of assessing process effectiveness microbiologically. However, there is a counter-balance to these positive attributes, brought about by variability in sterilisation process development and validation, and variability in the resistance properties (D-values) of the BIs. A change in D-value occurring between process validation and re-validation may result in different probabilities of BIs surviving on the second or subsequent validations or revalidations. Possible reasons behind such BI variability are discussed and potential ways forward proposed.

Corresponding author: Dr DG Allison, School of Pharmacy and Pharmaceutical Sciences, University of Manchester, Manchester M13 9PL, UK. Tel: +44 (0)161 275 2359, fax: +44 (0)161 275 2396, email: dallison@man.ac.uk

Sterilisation of biotechnology products

Robert A Pietrowski

Partner, David Begg Associates, Kirkbymoorside, North Yorkshire, UK

Biotechnology products are a heterogeneous group of pharmaceuticals derived from prokaryotic or eukaryotic cell culture and are typically heat-labile peptides, proteins or nucleic acids. As such, they are not amenable to terminal sterilisation by heat or other processes which damage biopolymers. Biotechnology products are, therefore, almost without exception, sterilised by submicron filtration and filled and finished using aseptic processing. The physical and chemical attributes of biotechnology products present a special challenge to the filtration process, demanding stringent control if product safety is to be assured.

The potential applicability of other means of sterilisation, such as nanofiltration and high-intensity, broad-spectrum pulsed light is critically discussed. Whilst these techniques may one day improve the assurance of sterility of this important group of products, their current utility is, at best, limited.

Corresponding author: Robert A Pietrowski, Partner, David Begg Associates, Kirkbymoorside, North Yorkshire, YO62 6AF, UK.

Tel: +44 (0)1751 432 999; fax: +44 (0)1751 432450;

email: mail@david-begg-associates.com

Coping with leaks in pharmaceutical isolators

Pharmaceutical isolators: protection for the product, the operator or both

Brian Midcalf

Consultant Pharmacist, Department of Continuing Education, University of Leeds, UK

Using an isolator may give assurance that a hazardous material, which is to be manipulated, will be separated completely from its manipulator. Further consideration of the isolator solution leads to a number of questions or options that need further consideration. Although this paper concentrates on the issues of handling potent or hazardous materials, isolators are used in many different environments or for different procedures.

An isolator may have a number of different transfer devices and can be constructed with a range of different features, so it is important that the optimum configuration is wisely selected. If we accept the principle that, in

practice, an isolator will leak, and that the substances to be handled are toxic or hazardous, it is likely to be a source of anxiety for the operator. This problem has been raised with the Medicines Control Agency (MCA) and the Health and Safety Executive (HSE) and a preliminary study has been completed. This study looked at the contamination levels of a cytotoxic agent in the workers using either a positive or negative pressure isolator. A survey of the use of isolators has been carried out through the National Quality Control Committee. It is important to keep the perception of leaks in perspective: the safe consideration of the product and the operator may be dependent on this. It is possible to estimate a leak quantitatively and these values could aid the safe use of an isolator.

The text for this paper is based on part of a presentation made at the Parenteral Society Conference, October 2000.

Corresponding author: Mr Brian Midcalf, 23 Riverside Drive, Otley, Yorkshire, LS21 2RU, UK. Tel: +44 (0)1943 464632; email: brian@midcalf.club24.co.uk